


Individual Report**QCMD 2020 Coronavirus Outbreak Preparedness EQA Pilot Study**

Catalogue Code: QAV204214	Ref Code: CVOP20	Challenge: S	Analysis Type: Qualitative	Dataset: 348360	Report UID: 11181/348360/2667	Laboratory IT282
-------------------------------------	----------------------------	------------------------	--------------------------------------	---------------------------	---	----------------------------

Intended Results / Panel Composition

Sample Code	Sample Content	Matrix	Sample Relationships ^[1]	Detection Frequency ^[2]	Sample Status ^[3]	Percentage Correct (All) ^[4]	
						(%)	(n)
CVOP20S-01	SARS-CoV-2	TM	4.30 dPCR Log10 Copies/ml (D1, DS1_2)	Frequently Detected	CORE	98.1	521
CVOP20S-02	Coronavirus NL63	TM	4.64 dPCR Log10 Copies/ml	Negative	EDUCATIONAL	96.9	521
CVOP20S-03	SARS-CoV-2	TM	3.30 dPCR Log10 Copies/ml (DS1_3)	Frequently Detected	CORE	96.9	521
CVOP20S-04	Coronavirus OC43	TM	4.03 dPCR Log10 Copies/ml	Negative	EDUCATIONAL	97.1	521
CVOP20S-05	Negative	TM		Negative	CORE	97.3	521
CVOP20S-06	SARS-CoV-2	TM	4.30 dPCR Log10 Copies/ml (D1, DS1_2)	Frequently Detected	CORE	98.5	521
CVOP20S-07	SARS-CoV-2	TM	5.30 dPCR Log10 Copies/ml (DS1_1)	Frequently Detected	CORE	99.2	521
CVOP20S-08	SARS-CoV-2	TM	2.30 dPCR Log10 Copies/ml (DS1_4)	Detected	CORE	86.0	521

Individual Report	QCMD 2020 Coronavirus Outbreak Preparedness EQA Pilot Study					
Catalogue Code: QAV204214	Ref Code: CVOP20	Challenge: S	Analysis Type: Qualitative	Dataset: 348360	Report UID: 11181/348360/2667	Laboratory IT282

[1] **Sample Relationships:** Indicates the relationships of the samples within this challenge. The highest titre member of dilution series DS1 is indicated by DS1_1 and further members of the series as DS1_2, DS1_3 etc. in order of reducing titre. Additional dilution series are indicated by DS2 (e.g DS2_1, DS2_2 etc.), DS3 (e.g. DS3_1, DS3_2 etc.). If one duplicate pair is present this is indicated by 'D1'. Further duplicate pairs are indicated by 'D2', 'D3' etc.

[2] **Detection Frequency:** To aid qualitative analysis each panel member is assigned a frequency of detection. This is based on the peer group consensus of all qualitative results returned from participants within the EQA challenge / distribution.

[3] **Sample Status:** EQA samples are defined as "CORE" or "EDUCATIONAL". Core proficiency samples are reviewed by the QCMD Scientific Expert(s). This is on the basis of scientific information, clinical relevance, current literature and, where appropriate, professional clinical guidelines. Participating laboratories are expected to report core proficiency samples correctly within the EQA challenge / distribution.

[4] **Percentage Correct (All):** Percentage of datasets (%) reporting the correct qualitative result and the total number of datasets (n) reported for each panel member.

For further details please refer to the current participant manual.

[1] **Sample Relationships:** Includes dPCR Log10 Copies/ml, the value obtained using a digital droplet PCR assay (modified from Eurosurveillance Jan 2020 Corman et al).

Please note: The values provided are for reference only. The values obtained by individual participating laboratories will vary from the dPCR values provided. This is because quantitation is dependent on the molecular workflow each laboratory uses (i.e. nucleic acid extraction and molecular platform / assay) as well as the standards or reference material used to calibrate the molecular workflow. In the absence of an International Standard or Certified Reference Material (CRM), QCMD uses Internal Reference Materials (IRMs) with values established using a dPCR reference assay in order to support the consistency and traceability of the EQA materials. This also helps aid the comparison of results across laboratories.


Samples **CVOP20S-02** and **CVOP20S-04** were included in the panel as educational specificity samples. For this assessment, these samples have been assigned as SARS-CoV-2 Negative.

Your Summary Results

EQA Assessment Group ^[1]	Seegene Allplex SARS-CoV-2
Core Panel Detection (Qualitative) Score ^[2]	<input type="text" value="0"/>

Core Panel Members Results

Sample Code	Qualitative Results			Your Quantitative Data (for information only) ^[3]		
	Percentage Correct (All) ^[4]	Your Result ^[5]	Detection Score ^[6]	Reported Value	Unitage	Cycle Threshold
CVOP20S-01	98.1	Positive	0		N/A	27.78
CVOP20S-03	96.9	Positive	0		N/A	31.27
CVOP20S-05	97.3	Negative	0		N/A	27.84
CVOP20S-06	98.5	Positive	0		N/A	27.82
CVOP20S-07	99.2	Positive	0		N/A	27.46
CVOP20S-08	86.0	Positive	0		N/A	35.21

Individual Report	QCMD 2020 Coronavirus Outbreak Preparedness EQA Pilot Study				 <small>Quality Control for Molecular Diagnostics</small>	
Catalogue Code: QAV204214	Ref Code: CVOP20	Challenge: S	Analysis Type: Qualitative	Dataset: 348360	Report UID: 11181/348360/2667	Laboratory IT282

[1] **EQA Assessment Group:** To aid data analysis, participant results are grouped according to the molecular amplification/detection method specified within their molecular workflow for this challenge / distribution. For further details refer to the *Additional Information: Individual Panel Member Analysis* section of this report.

[2] **Core Panel Detection (Qualitative) Score:** An overall core panel detection score provided per challenge / distribution.

[3] **Quantitative Data (for information only):** This is the quantitative value, unitage and cycle threshold you provided when you submitted your results. For qualitative programmes this information is not used as part of your formal EQA assessment.


[4] **Percentage Correct (All):** Percentage of datasets (%) reporting the correct qualitative results for each panel member.

[5] **Your Result:** The qualitative result you reported for each sample within this EQA challenge / distribution.

[6] **Detection Score:** Your detection (qualitative) scores are based on the assigned detection frequency of each panel members, where 0 (zero) is "highly satisfactory" and 3 (three) is "highly unsatisfactory". Scores are provided for individual panel members.


For further details please refer to the current participant manual.

Core Panel Member Score Breakdown






Core Panel Member Score Breakdown - Detection: This figure gives you a breakdown of the qualitative detection scores for all qualitative datasets returned within this EQA challenge / distribution independent of the EQA assessment group. Panel detection scores are generated from only those panel members that are defined as "CORE".

For further details please refer to the current participant manual.

Individual Report	QCMD 2020 Coronavirus Outbreak Preparedness EQA Pilot Study				 <small>Quality Control for Molecular Diagnostics</small>	
Catalogue Code: QAV204214	Ref Code: CVOP20	Challenge: S	Analysis Type: Qualitative	Dataset: 348360	Report UID: 11181/348360/2667	Laboratory IT282

My Workflow Details

The details of the workflow(s) used to submit your results for this challenge.

Name	Coronavirus SARS-2 (v3)
Description	Coronavirus SARS-2 test: RNA extraction by QIAAsymphony DSP Virus/Pathogen Kit - QIAGEN (default IC 400 uL programme - elution in 60 uL); amplification by Real time PCR Allplex™ 2019 n-CoV Assay, Seegene.
Targets	 coronavirus
Assays	<ul style="list-style-type: none">  <i>Extraction</i> - QIAGEN - QIAAsymphony SP System <ul style="list-style-type: none"> ● Commercial <ul style="list-style-type: none"> ○ Kit Manufacturer: <i>QIAGEN</i> ○ Kit Type: <i>QIAAsymphony DSP Virus/Pathogen Midi Kit</i>  <i>Amplification</i> - Bio-Rad - CFX96 Touch Deep Well Real-Time PCR Detection System <ul style="list-style-type: none"> ● Multiplex ● Commercial <ul style="list-style-type: none"> ○ Kit Manufacturer: <i>Seegene</i> ○ Kit Type: <i>Allplex™ 2019 n-CoV Assay</i> ○ Kit Version: <i>Allplex™ 2019 n-CoV Assay</i>

Educational Panel Members Results

Sample Code	Qualitative Results			Your Quantitative Data (for information only) ^[1]		
	Percentage Correct (All) ^[2]	Your Result ^[3]	Detection Score ^[4]	Reported Value	Unitage	Cycle Threshold
CVOP20S-02	96.9	Negative	0		N/A	30.27
CVOP20S-04	97.1	Negative	0		N/A	36.67


[1] Quantitative Data (for information only): This is the quantitative value, unitage and cycle threshold you provided when you submitted your results. For qualitative programmes this information is not used as part of your formal EQA assessment.

[2] Percentage Correct (All): Percentage of datasets (%) reporting the correct qualitative results for each panel member.

[3] Your Result: The qualitative result you reported for each sample within this EQA challenge / distribution.

[4] Detection Score: Your detection (qualitative) scores are based on the assigned detection frequency of each panel members, where 0 (zero) is "highly satisfactory" and 3 (three) is "highly unsatisfactory". Scores are provided for individual panel members.

For further details please refer to the current participant manual.

Individual Report	QCMD 2020 Coronavirus Outbreak Preparedness EQA Pilot Study					
Catalogue Code: QAV204214	Ref Code: CVOP20	Challenge: S	Analysis Type: Qualitative	Dataset: 348360	Report UID: 11181/348360/2667	Laboratory IT282

Further Programme Details

Number of Participants	406
Number of Countries	38
Number of Respondents	365
Number of Datasets Submitted	521
Qualitative Results Returned	521 (100.0%)

Comments

Thirty-six (7.1%) laboratories reported sample CVOP20S-02 as Negative for SARS-CoV-2 and correctly reported the sample as Coronavirus NL63.

Thirty-seven (7.2%) laboratories reported sample CVOP20S-04 as Negative for SARS-CoV-2 and correctly reported the sample as Coronavirus OC43.


EQA Programme Aims

To assess the proficiency of laboratories molecular technologies for the detection and determination of SARS-CoV-2 from other coronaviruses.

Feedback and Enquiries

Participants are encouraged to read the QCMD Participants' Manual, which can be downloaded from the QCMD website.

Any enquiries should be submitted through the 'Contact Us' form that you can find in the 'Help' section of your QCMD (ITEMS) Participant Profile Area.

Individual Report	QCMD 2020 Coronavirus Outbreak Preparedness EQA Pilot Study					
Catalogue Code: QAV204214	Ref Code: CVOP20	Challenge: S	Analysis Type: Qualitative	Dataset: 348360	Report UID: 11181/348360/2667	Laboratory IT282

Panel member analysis is separated into CORE samples followed by EDUCATIONAL samples.

Additional Core Samples Information

The following section has been categorised as shown below:

Core ► Qualitative

Individual Panel Member Analysis (Qualitative)


Qualitative analysis for each panel member is provided in relation to your EQA assessment group. EQA assessment groups are established using the molecular workflow information reported by all participants within this EQA challenge / distribution. The principal level of assessment is at the individual method level which is defined based on your reported “amplification/detection method” and other laboratories using the same or similar amplification/detection methods.


To allow meaningful assessment at the individual method level the EQA assessment group must consist of 5 or more datasets. If there are not sufficient datasets at the individual method level then your results will be included within a higher EQA assessment group based on whether it is a commercial or in house technology/method. The highest level assessment grouping is “All” participant reported qualitative results.

A breakdown of qualitative results reported by participants on each of the panel members within this EQA challenge / distribution is provided below. You can compare your results to those within your EQA assessment group and those obtained within other EQA assessment groups or to the overall consensus for each sample within this EQA challenge / distribution.

CVOP20S-01 - Qualitative Results Breakdown

Sample Code	Sample Content	Matrix	Sample Relationships	Detection Frequency	Sample Status	Percentage Correct (All)	
						(%)	(n)
CVOP20S-01	SARS-CoV-2	TM	4.30 dPCR Log10 Copies/ml (D1, DS1_2)	Frequently Detected	CORE	98.1	521




Individual Report	QCMD 2020 Coronavirus Outbreak Preparedness EQA Pilot Study					
	Catalogue Code: QAV204214	Ref Code: CVOP20	Challenge: S	Analysis Type: Qualitative	Dataset: 348360	Report UID: 11181/348360/2667


Groups below n=5: AB Analytica (n=3), AB Analytica - AB Analytica REALQUALITY RQ (n=3), ABL (n=1), ABL - ABL UltraGene (n=1), Anchor Diagnostics (n=3), Anchor Diagnostics - Anchor Diagnostics SARS CoV-2 (n=3), Applied Biosystems (n=2), Applied Biosystems - Applied Biosystems Real Time PCR Kit (n=2), Appolon biotek (n=1), Appolon biotek - Appolon biotek SARS-CoV-2 (n=1), AusDiagnostics - AusDiagnostics Easy-Plex (n=4), BGI (n=2), BGI - BGI Real Time PCR (n=2), BioGX (n=1), BioGX - BioGX SARS-CoV-2 (n=1), Bioeksen (n=2), Bioeksen - Bioeksen Biospeedy SARS-CoV-2 (n=2), CDC (n=3), CDC - CDC SARS-CoV-2 (n=3), Co-Diagnostics Inc (n=3), Co-Diagnostics Inc - Co-Diagnostics Inc Logix Smart (n=3), DiaCarta (n=1), DiaCarta - DiaCarta QuantiVirus PCR kit (n=1), DiaSorin (n=2), DiaSorin - DiaSorin Simplexa (n=2), Diagenode (n=1), Diagenode - Diagenode Real-Time PCR (n=1), EUROIMMUN (n=3), EUROIMMUN - EUROIMMUN SARS-CoV-2 (n=3), GenMark (n=1), GenMark - GenMark SARS-CoV-2 (n=1), GeneProof (n=4), GeneProof - GeneProof SARS-CoV-2 (n=4), Genematrix (n=1), Genematrix - Genematrix NeoPlex (n=1), Hologic (n=2), Hologic - Hologic Panther Fusion (n=2), HybriBio Limited (n=1), HybriBio Limited - HybriBio Real-Time PCR Kit (n=1), Immundiagnostik (n=1), Immundiagnostik - Immundiagnostik MutaPLEX (n=1), Kogene Biotech (n=1), Kogene Biotech - Kogene Powerchek (n=1), Liferiver (n=1), Liferiver - Liferiver 2019-nCoV (n=1), Luminex - Luminex ARIES (n=2), Luminex - Luminex xTAG (n=4), Mediven (n=1), Mediven - Mediven GenoAmp PCR (n=1), Mikrogen (n=4), Mikrogen - Mikrogen ampliCube (n=4), Mobidiag (n=4), Mobidiag - Mobidiag Amplidiag (n=4), QIAGEN (n=4), QIAGEN - Qiagen QIAstat-Dx (n=4), Randox (n=2), Randox - Randox COVID-19 Array (n=1), Randox - Randox Extended Array (n=1), Sansure Biotech (n=1), Sansure Biotech - Sansure Biotech nCoV PCR Kit (n=1), SeroSep (n=1), SeroSep - SeroSep Real Time PCR (n=1), Shanghai BioGerm (n=1), Shanghai BioGerm - Shanghai BioGerm Real Time PCR (n=1), Shanghai Fosun Pharmaceutical (n=1), Shanghai Fosun Pharmaceutical - Shanghai Fosun Real Time PCR (n=1), Vazyme (n=1), Vazyme - Vazyme SARS-CoV-2 (n=1), Vitassay (n=4), Vitassay - Vitassay SARS-CoV-2 (n=4), bioMerieux (n=4), bioMerieux - BioFire FilmArray (n=1), bioMerieux - bioMerieux R-gene SARS-CoV-2 (n=3), fast-track DIAGNOSTICS (n=2), fast-track DIAGNOSTICS - FTD real time PCR (n=2)

Groups Rolled Up: Abbott - Abbott SARS-CoV-2 (n=17), Altona Diagnostics - Altona Diagnostics RealStar (n=28), Anatolia Geneworks - Anatolia Geneworks Bosphore nCoV (n=9), Cepheid - Cepheid Xpert SARS-CoV-2 (n=51), ELITech Group - Elitech GeneFinder (n=25), PrimerDesign - PrimerDesign Genesig (n=7), R-Biopharm - R-Biopharm RIDA Gene (n=5), Roche - Roche Cobas SARS-CoV-2 (n=37), TIB MOLBIOL - TIB MOLBIOL LightMix (n=27), ThermoFisher - ThermoFisher TaqPath COVID-19 (n=6), In-House - Real-time In-House PCR (n=119)

CVOP20S-03 - Qualitative Results Breakdown


Sample Code	Sample Content	Matrix	Sample Relationships	Detection Frequency	Sample Status	Percentage Correct (All)	
						(%)	(n)
CVOP20S-03	SARS-CoV-2	TM	3.30 dPCR Log10 Copies/ml (DS1_3)	Frequently Detected	CORE	96.9	521



Individual Report	QCMD 2020 Coronavirus Outbreak Preparedness EQA Pilot Study				 <small>Quality Control for Molecular Diagnostics</small>	
	Catalogue Code: QAV204214	Ref Code: CVOP20	Challenge: S	Analysis Type: Qualitative	Dataset: 348360	Report UID: 11181/348360/2667


Groups below n=5: AB Analytica (n=3), AB Analytica - AB Analytica REALQUALITY RQ (n=3), ABL (n=1), ABL - ABL UltraGene (n=1), Anchor Diagnostics (n=3), Anchor Diagnostics - Anchor Diagnostics SARS CoV-2 (n=3), Applied Biosystems (n=2), Applied Biosystems - Applied Biosystems Real Time PCR Kit (n=2), Appolon biotek (n=1), Appolon biotek - Appolon biotek SARS-CoV-2 (n=1), AusDiagnostics - AusDiagnostics Easy-Plex (n=4), BGI (n=2), BGI - BGI Real Time PCR (n=2), BioGX (n=1), BioGX - BioGX SARS-CoV-2 (n=1), Bioeksen (n=2), Bioeksen - Bioeksen Biospeedy SARS-CoV-2 (n=2), CDC (n=3), CDC - CDC SARS-CoV-2 (n=3), Co-Diagnostics Inc (n=3), Co-Diagnostics Inc - Co-Diagnostics Inc Logix Smart (n=3), DiaCarta (n=1), DiaCarta - DiaCarta QuantiVirus PCR kit (n=1), DiaSorin (n=2), DiaSorin - DiaSorin Simplexa (n=2), Diagenode (n=1), Diagenode - Diagenode Real-Time PCR (n=1), EUROIMMUN (n=3), EUROIMMUN - EUROIMMUN SARS-CoV-2 (n=3), GenMark (n=1), GenMark - GenMark SARS-CoV-2 (n=1), GeneProof (n=4), GeneProof - GeneProof SARS-CoV-2 (n=4), Genematrix (n=1), Genematrix - Genematrix NeoPlex (n=1), Hologic (n=2), Hologic - Hologic Panther Fusion (n=2), HybriBio Limited (n=1), HybriBio Limited - HybriBio Real-Time PCR Kit (n=1), Immundiagnostik (n=1), Immundiagnostik - Immundiagnostik MutaPLEX (n=1), Kogene Biotech (n=1), Kogene Biotech - Kogene Powerchek (n=1), Liferiver (n=1), Liferiver - Liferiver 2019-nCoV (n=1), Luminex - Luminex ARIES (n=2), Luminex - Luminex xTAG (n=4), Mediven (n=1), Mediven - Mediven GenoAmp PCR (n=1), Mikrogen (n=4), Mikrogen - Mikrogen ampliCube (n=4), Mobidiag (n=4), Mobidiag - Mobidiag Amplidiag (n=4), QIAGEN (n=4), QIAGEN - Qiagen QIAstat-Dx (n=4), Randox (n=2), Randox - Randox COVID-19 Array (n=1), Randox - Randox Extended Array (n=1), Sansure Biotech (n=1), Sansure Biotech - Sansure Biotech nCoV PCR Kit (n=1), SeroSep (n=1), SeroSep - SeroSep Real Time PCR (n=1), Shanghai BioGerm (n=1), Shanghai BioGerm - Shanghai BioGerm Real Time PCR (n=1), Shanghai Fosun Pharmaceutical (n=1), Shanghai Fosun Pharmaceutical - Shanghai Fosun Real Time PCR (n=1), Vazyme (n=1), Vazyme - Vazyme SARS-CoV-2 (n=1), Vitassay (n=4), Vitassay - Vitassay SARS-CoV-2 (n=4), bioMerieux (n=4), bioMerieux - BioFire FilmArray (n=1), bioMerieux - bioMerieux R-gene SARS-CoV-2 (n=3), fast-track DIAGNOSTICS (n=2), fast-track DIAGNOSTICS - FTD real time PCR (n=2)


Groups Rolled Up: Abbott - Abbott SARS-CoV-2 (n=17), Altona Diagnostics - Altona Diagnostics RealStar (n=28), Anatolia Geneworks - Anatolia Geneworks Bosphore nCoV (n=9), Cepheid - Cepheid Xpert SARS-CoV-2 (n=51), ELITech Group - Elitech GeneFinder (n=25), PrimerDesign - PrimerDesign Genesig (n=7), R-Biopharm - R-Biopharm RIDA Gene (n=5), Roche - Roche Cobas SARS-CoV-2 (n=37), TIB MOLBIOL - TIB MOLBIOL LightMix (n=27), ThermoFisher - ThermoFisher TaqPath COVID-19 (n=6), In-House - Real-time In-House PCR (n=119)

Individual Report	QCMD 2020 Coronavirus Outbreak Preparedness EQA Pilot Study				 <small>Quality Control for Molecular Diagnostics</small>	
Catalogue Code: QAV204214	Ref Code: CVOP20	Challenge: S	Analysis Type: Qualitative	Dataset: 348360	Report UID: 11181/348360/2667	Laboratory IT282

CVOP20S-05 - Qualitative Results Breakdown

Sample Code	Sample Content	Matrix	Sample Relationships	Detection Frequency	Sample Status	Percentage Correct (All)	
						(%)	(n)
CVOP20S-05	Negative	TM		Negative	CORE	97.3	521




Individual Report	QCMD 2020 Coronavirus Outbreak Preparedness EQA Pilot Study					
	Catalogue Code: QAV204214	Ref Code: CVOP20	Challenge: S	Analysis Type: Qualitative	Dataset: 348360	Report UID: 11181/348360/2667


Groups below n=5: AB Analytica (n=3), AB Analytica - AB Analytica REALQUALITY RQ (n=3), ABL (n=1), ABL - ABL UltraGene (n=1), Anchor Diagnostics (n=3), Anchor Diagnostics - Anchor Diagnostics SARS CoV-2 (n=3), Applied Biosystems (n=2), Applied Biosystems - Applied Biosystems Real Time PCR Kit (n=2), Appolon biotek (n=1), Appolon biotek - Appolon biotek SARS-CoV-2 (n=1), AusDiagnostics - AusDiagnostics Easy-Plex (n=4), BGI (n=2), BGI - BGI Real Time PCR (n=2), BioGX (n=1), BioGX - BioGX SARS-CoV-2 (n=1), Bioeksen (n=2), Bioeksen - Bioeksen Biospeedy SARS-CoV-2 (n=2), CDC (n=3), CDC - CDC SARS-CoV-2 (n=3), Co-Diagnostics Inc (n=3), Co-Diagnostics Inc - Co-Diagnostics Inc Logix Smart (n=3), DiaCarta (n=1), DiaCarta - DiaCarta QuantiVirus PCR kit (n=1), DiaSorin (n=2), DiaSorin - DiaSorin Simplexa (n=2), Diagenode (n=1), Diagenode - Diagenode Real-Time PCR (n=1), EUROIMMUN (n=3), EUROIMMUN - EUROIMMUN SARS-CoV-2 (n=3), GenMark (n=1), GenMark - GenMark SARS-CoV-2 (n=1), GeneProof (n=4), GeneProof - GeneProof SARS-CoV-2 (n=4), Genematrix (n=1), Genematrix - Genematrix NeoPlex (n=1), Hologic (n=2), Hologic - Hologic Panther Fusion (n=2), HybriBio Limited (n=1), HybriBio Limited - HybriBio Real-Time PCR Kit (n=1), Immundiagnostik (n=1), Immundiagnostik - Immundiagnostik MutaPLEX (n=1), Kogene Biotech (n=1), Kogene Biotech - Kogene Powerchek (n=1), Liferiver (n=1), Liferiver - Liferiver 2019-nCoV (n=1), Luminex - Luminex ARIES (n=2), Luminex - Luminex xTAG (n=4), Mediven (n=1), Mediven - Mediven GenoAmp PCR (n=1), Mikrogen (n=4), Mikrogen - Mikrogen ampliCube (n=4), Mobidiag (n=4), Mobidiag - Mobidiag Amplidiag (n=4), QIAGEN (n=4), QIAGEN - Qiagen QIAstat-Dx (n=4), Randox (n=2), Randox - Randox COVID-19 Array (n=1), Randox - Randox Extended Array (n=1), Sansure Biotech (n=1), Sansure Biotech - Sansure Biotech nCoV PCR Kit (n=1), SeroSep (n=1), SeroSep - SeroSep Real Time PCR (n=1), Shanghai BioGerm (n=1), Shanghai BioGerm - Shanghai BioGerm Real Time PCR (n=1), Shanghai Fosun Pharmaceutical (n=1), Shanghai Fosun Pharmaceutical - Shanghai Fosun Real Time PCR (n=1), Vazyme (n=1), Vazyme - Vazyme SARS-CoV-2 (n=1), Vitassay (n=4), Vitassay - Vitassay SARS-CoV-2 (n=4), bioMerieux (n=4), bioMerieux - BioFire FilmArray (n=1), bioMerieux - bioMerieux R-gene SARS-CoV-2 (n=3), fast-track DIAGNOSTICS (n=2), fast-track DIAGNOSTICS - FTD real time PCR (n=2)

Groups Rolled Up: Abbott - Abbott SARS-CoV-2 (n=17), Altona Diagnostics - Altona Diagnostics RealStar (n=28), Anatolia Geneworks - Anatolia Geneworks Bosphore nCoV (n=9), Cepheid - Cepheid Xpert SARS-CoV-2 (n=51), ELITech Group - Elitech GeneFinder (n=25), PrimerDesign - PrimerDesign Genesig (n=7), R-Biopharm - R-Biopharm RIDA Gene (n=5), Roche - Roche Cobas SARS-CoV-2 (n=37), TIB MOLBIOL - TIB MOLBIOL LightMix (n=27), ThermoFisher - ThermoFisher TaqPath COVID-19 (n=6), In-House - Real-time In-House PCR (n=119)

CVOP20S-06 - Qualitative Results Breakdown

Sample Code	Sample Content	Matrix	Sample Relationships	Detection Frequency	Sample Status	Percentage Correct (All)	
						(%)	(n)
CVOP20S-06	SARS-CoV-2	TM	4.30 dPCR Log10 Copies/ml (D1, DS1_2)	Frequently Detected	CORE	98.5	521




Individual Report	QCMD 2020 Coronavirus Outbreak Preparedness EQA Pilot Study					
	Catalogue Code: QAV204214	Ref Code: CVOP20	Challenge: S	Analysis Type: Qualitative	Dataset: 348360	Report UID: 11181/348360/2667


Groups below n=5: AB Analytica (n=3), AB Analytica - AB Analytica REALQUALITY RQ (n=3), ABL (n=1), ABL - ABL UltraGene (n=1), Anchor Diagnostics (n=3), Anchor Diagnostics - Anchor Diagnostics SARS CoV-2 (n=3), Applied Biosystems (n=2), Applied Biosystems - Applied Biosystems Real Time PCR Kit (n=2), Appolon biotek (n=1), Appolon biotek - Appolon biotek SARS-CoV-2 (n=1), AusDiagnostics - AusDiagnostics Easy-Plex (n=4), BGI (n=2), BGI - BGI Real Time PCR (n=2), BioGX (n=1), BioGX - BioGX SARS-CoV-2 (n=1), Bioeksen (n=2), Bioeksen - Bioeksen Biospeedy SARS-CoV-2 (n=2), CDC (n=3), CDC - CDC SARS-CoV-2 (n=3), Co-Diagnostics Inc (n=3), Co-Diagnostics Inc - Co-Diagnostics Inc Logix Smart (n=3), DiaCarta (n=1), DiaCarta - DiaCarta QuantiVirus PCR kit (n=1), DiaSorin (n=2), DiaSorin - DiaSorin Simplexa (n=2), Diagenode (n=1), Diagenode - Diagenode Real-Time PCR (n=1), EUROIMMUN (n=3), EUROIMMUN - EUROIMMUN SARS-CoV-2 (n=3), GenMark (n=1), GenMark - GenMark SARS-CoV-2 (n=1), GeneProof (n=4), GeneProof - GeneProof SARS-CoV-2 (n=4), Genematrix (n=1), Genematrix - Genematrix NeoPlex (n=1), Hologic (n=2), Hologic - Hologic Panther Fusion (n=2), HybriBio Limited (n=1), HybriBio Limited - HybriBio Real-Time PCR Kit (n=1), Immundiagnostik (n=1), Immundiagnostik - Immundiagnostik MutaPLEX (n=1), Kogene Biotech (n=1), Kogene Biotech - Kogene Powerchek (n=1), Liferiver (n=1), Liferiver - Liferiver 2019-nCoV (n=1), Luminex - Luminex ARIES (n=2), Luminex - Luminex xTAG (n=4), Mediven (n=1), Mediven - Mediven GenoAmp PCR (n=1), Mikrogen (n=4), Mikrogen - Mikrogen ampliCube (n=4), Mobidiag (n=4), Mobidiag - Mobidiag Amplidiag (n=4), QIAGEN (n=4), QIAGEN - Qiagen QIAstat-Dx (n=4), Randox (n=2), Randox - Randox COVID-19 Array (n=1), Randox - Randox Extended Array (n=1), Sansure Biotech (n=1), Sansure Biotech - Sansure Biotech nCoV PCR Kit (n=1), SeroSep (n=1), SeroSep - SeroSep Real Time PCR (n=1), Shanghai BioGerm (n=1), Shanghai BioGerm - Shanghai BioGerm Real Time PCR (n=1), Shanghai Fosun Pharmaceutical (n=1), Shanghai Fosun Pharmaceutical - Shanghai Fosun Real Time PCR (n=1), Vazyme (n=1), Vazyme - Vazyme SARS-CoV-2 (n=1), Vitassay (n=4), Vitassay - Vitassay SARS-CoV-2 (n=4), bioMerieux (n=4), bioMerieux - BioFire FilmArray (n=1), bioMerieux - bioMerieux R-gene SARS-CoV-2 (n=3), fast-track DIAGNOSTICS (n=2), fast-track DIAGNOSTICS - FTD real time PCR (n=2)

Groups Rolled Up: Abbott - Abbott SARS-CoV-2 (n=17), Altona Diagnostics - Altona Diagnostics RealStar (n=28), Anatolia Geneworks - Anatolia Geneworks Bosphore nCoV (n=9), Cepheid - Cepheid Xpert SARS-CoV-2 (n=51), ELITech Group - Elitech GeneFinder (n=25), PrimerDesign - PrimerDesign Genesig (n=7), R-Biopharm - R-Biopharm RIDA Gene (n=5), Roche - Roche Cobas SARS-CoV-2 (n=37), TIB MOLBIOL - TIB MOLBIOL LightMix (n=27), ThermoFisher - ThermoFisher TaqPath COVID-19 (n=6), In-House - Real-time In-House PCR (n=119)

CVOP20S-07 - Qualitative Results Breakdown

Sample Code	Sample Content	Matrix	Sample Relationships	Detection Frequency	Sample Status	Percentage Correct (All)	
						(%)	(n)
CVOP20S-07	SARS-CoV-2	TM	5.30 dPCR Log10 Copies/ml (DS1_1)	Frequently Detected	CORE	99.2	521




Individual Report	QCMD 2020 Coronavirus Outbreak Preparedness EQA Pilot Study					
	Catalogue Code: QAV204214	Ref Code: CVOP20	Challenge: S	Analysis Type: Qualitative	Dataset: 348360	Report UID: 11181/348360/2667


Groups below n=5: AB Analytica (n=3), AB Analytica - AB Analytica REALQUALITY RQ (n=3), ABL (n=1), ABL - ABL UltraGene (n=1), Anchor Diagnostics (n=3), Anchor Diagnostics - Anchor Diagnostics SARS CoV-2 (n=3), Applied Biosystems (n=2), Applied Biosystems - Applied Biosystems Real Time PCR Kit (n=2), Appolon biotek (n=1), Appolon biotek - Appolon biotek SARS-CoV-2 (n=1), AusDiagnostics - AusDiagnostics Easy-Plex (n=4), BGI (n=2), BGI - BGI Real Time PCR (n=2), BioGX (n=1), BioGX - BioGX SARS-CoV-2 (n=1), Bioeksen (n=2), Bioeksen - Bioeksen Biospeedy SARS-CoV-2 (n=2), CDC (n=3), CDC - CDC SARS-CoV-2 (n=3), Co-Diagnostics Inc (n=3), Co-Diagnostics Inc - Co-Diagnostics Inc Logix Smart (n=3), DiaCarta (n=1), DiaCarta - DiaCarta QuantiVirus PCR kit (n=1), DiaSorin (n=2), DiaSorin - DiaSorin Simplexa (n=2), Diagenode (n=1), Diagenode - Diagenode Real-Time PCR (n=1), EUROIMMUN (n=3), EUROIMMUN - EUROIMMUN SARS-CoV-2 (n=3), GenMark (n=1), GenMark - GenMark SARS-CoV-2 (n=1), GeneProof (n=4), GeneProof - GeneProof SARS-CoV-2 (n=4), Genematrix (n=1), Genematrix - Genematrix NeoPlex (n=1), Hologic (n=2), Hologic - Hologic Panther Fusion (n=2), HybriBio Limited (n=1), HybriBio Limited - HybriBio Real-Time PCR Kit (n=1), Immundiagnostik (n=1), Immundiagnostik - Immundiagnostik MutaPLEX (n=1), Kogene Biotech (n=1), Kogene Biotech - Kogene Powerchek (n=1), Liferiver (n=1), Liferiver - Liferiver 2019-nCoV (n=1), Luminex - Luminex ARIES (n=2), Luminex - Luminex xTAG (n=4), Mediven (n=1), Mediven - Mediven GenoAmp PCR (n=1), Mikrogen (n=4), Mikrogen - Mikrogen ampliCube (n=4), Mobidiag (n=4), Mobidiag - Mobidiag Amplidiag (n=4), QIAGEN (n=4), QIAGEN - Qiagen QIAstat-Dx (n=4), Randox (n=2), Randox - Randox COVID-19 Array (n=1), Randox - Randox Extended Array (n=1), Sansure Biotech (n=1), Sansure Biotech - Sansure Biotech nCoV PCR Kit (n=1), SeroSep (n=1), SeroSep - SeroSep Real Time PCR (n=1), Shanghai BioGerm (n=1), Shanghai BioGerm - Shanghai BioGerm Real Time PCR (n=1), Shanghai Fosun Pharmaceutical (n=1), Shanghai Fosun Pharmaceutical - Shanghai Fosun Real Time PCR (n=1), Vazyme (n=1), Vazyme - Vazyme SARS-CoV-2 (n=1), Vitassay (n=4), Vitassay - Vitassay SARS-CoV-2 (n=4), bioMerieux (n=4), bioMerieux - BioFire FilmArray (n=1), bioMerieux - bioMerieux R-gene SARS-CoV-2 (n=3), fast-track DIAGNOSTICS (n=2), fast-track DIAGNOSTICS - FTD real time PCR (n=2)

Groups Rolled Up: Abbott - Abbott SARS-CoV-2 (n=17), Altona Diagnostics - Altona Diagnostics RealStar (n=28), Anatolia Geneworks - Anatolia Geneworks Bosphore nCoV (n=9), Cepheid - Cepheid Xpert SARS-CoV-2 (n=51), ELITech Group - Elitech GeneFinder (n=25), PrimerDesign - PrimerDesign Genesig (n=7), R-Biopharm - R-Biopharm RIDA Gene (n=5), Roche - Roche Cobas SARS-CoV-2 (n=37), TIB MOLBIOL - TIB MOLBIOL LightMix (n=27), ThermoFisher - ThermoFisher TaqPath COVID-19 (n=6), In-House - Real-time In-House PCR (n=119)

CVOP20S-08 - Qualitative Results Breakdown

Sample Code	Sample Content	Matrix	Sample Relationships	Detection Frequency	Sample Status	Percentage Correct (All)	
						(%)	(n)
CVOP20S-08	SARS-CoV-2	TM	2.30 dPCR Log10 Copies/ml (DS1_4)	Detected	CORE	86.0	521



Individual Report	QCMD 2020 Coronavirus Outbreak Preparedness EQA Pilot Study					
Catalogue Code: QAV204214	Ref Code: CVOP20	Challenge: S	Analysis Type: Qualitative	Dataset: 348360	Report UID: 11181/348360/2667	Laboratory IT282

Groups below n=5: AB Analytica (n=3), AB Analytica - AB Analytica REALQUALITY RQ (n=3), ABL (n=1), ABL - ABL UltraGene (n=1), Anchor Diagnostics (n=3), Anchor Diagnostics - Anchor Diagnostics SARS CoV-2 (n=3), Applied Biosystems (n=2), Applied Biosystems - Applied Biosystems Real Time PCR Kit (n=2), Appolon biotek (n=1), Appolon biotek - Appolon biotek SARS-CoV-2 (n=1), AusDiagnostics - AusDiagnostics Easy-Plex (n=4), BGI (n=2), BGI - BGI Real Time PCR (n=2), BioGX (n=1), BioGX - BioGX SARS-CoV-2 (n=1), Bioeksen (n=2), Bioeksen - Bioeksen Biospeedy SARS-CoV-2 (n=2), CDC (n=3), CDC - CDC SARS-CoV-2 (n=3), Co-Diagnostics Inc (n=3), Co-Diagnostics Inc - Co-Diagnostics Inc Logix Smart (n=3), DiaCarta (n=1), DiaCarta - DiaCarta QuantiVirus PCR kit (n=1), DiaSorin (n=2), DiaSorin - DiaSorin Simplexa (n=2), Diagenode (n=1), Diagenode - Diagenode Real-Time PCR (n=1), EUROIMMUN (n=3), EUROIMMUN - EUROIMMUN SARS-CoV-2 (n=3), GenMark (n=1), GenMark - GenMark SARS-CoV-2 (n=1), GeneProof (n=4), GeneProof - GeneProof SARS-CoV-2 (n=4), Genematrix (n=1), Genematrix - Genematrix NeoPlex (n=1), Hologic (n=2), Hologic - Hologic Panther Fusion (n=2), HybriBio Limited (n=1), HybriBio Limited - HybriBio Real-Time PCR Kit (n=1), Immundiagnostik (n=1), Immundiagnostik - Immundiagnostik MutaPLEX (n=1), Kogene Biotech (n=1), Kogene Biotech - Kogene Powerchek (n=1), Liferiver (n=1), Liferiver - Liferiver 2019-nCoV (n=1), Luminex - Luminex ARIES (n=2), Luminex - Luminex xTAG (n=4), Mediven (n=1), Mediven - Mediven GenoAmp PCR (n=1), Mikrogen (n=4), Mikrogen - Mikrogen ampliCube (n=4), Mobidiag (n=4), Mobidiag - Mobidiag Amplidiag (n=4), QIAGEN (n=4), QIAGEN - Qiagen QIAstat-Dx (n=4), Randox (n=2), Randox - Randox COVID-19 Array (n=1), Randox - Randox Extended Array (n=1), Sansure Biotech (n=1), Sansure Biotech - Sansure Biotech nCoV PCR Kit (n=1), Serosep (n=1), Serosep - Serosep Real Time PCR (n=1), Shanghai BioGerm (n=1), Shanghai BioGerm - Shanghai BioGerm Real Time PCR (n=1), Shanghai Fosun Pharmaceutical (n=1), Shanghai Fosun Pharmaceutical - Shanghai Fosun Real Time PCR (n=1), Vazyme (n=1), Vazyme - Vazyme SARS-CoV-2 (n=1), Vitassay (n=4), Vitassay - Vitassay SARS-CoV-2 (n=4), bioMerieux (n=4), bioMerieux - BioFire FilmArray (n=1), bioMerieux - bioMerieux R-gene SARS-CoV-2 (n=3), fast-track DIAGNOSTICS (n=2), fast-track DIAGNOSTICS - FTD real time PCR (n=2)

Groups Rolled Up: Abbott - Abbott SARS-CoV-2 (n=17), Altona Diagnostics - Altona Diagnostics RealStar (n=28), Anatolia Geneworks - Anatolia Geneworks Bosphore nCoV (n=9), Cepheid - Cepheid Xpert SARS-CoV-2 (n=51), ELITech Group - Elitech GeneFinder (n=25), PrimerDesign - PrimerDesign Genesig (n=7), R-Biopharm - R-Biopharm RIDA Gene (n=5), Roche - Roche Cobas SARS-CoV 2 (n=37), TIB MOLBIOL - TIB MOLBIOL LightMix (n=27), ThermoFisher - ThermoFisher TaqPath COVID-19 (n=6), In-House - Real-time In-House PCR (n=119)

Additional Educational Samples Information

The following section has been categorised as shown below:

Educational ► Qualitative

Individual Panel Member Analysis (Qualitative)


Qualitative analysis for each panel member is provided in relation to your EQA assessment group. EQA assessment groups are established using the molecular workflow information reported by all participants within this EQA challenge / distribution. The principal level of assessment is at the individual method level which is defined based on your reported “amplification/detection method” and other laboratories using the same or similar amplification/detection methods.


To allow meaningful assessment at the individual method level the EQA assessment group must consist of 5 or more datasets. If there are not sufficient datasets at the individual method level then your results will be included within a higher EQA assessment group based on whether it is a commercial or in house technology/method. The highest level assessment grouping is “All” participant reported qualitative results.

A breakdown of qualitative results reported by participants on each of the panel members within this EQA challenge / distribution is provided below. You can compare your results to those within your EQA assessment group and those obtained within other EQA assessment groups or to the overall consensus for each sample within this EQA challenge / distribution.

CVOP20S-02 - Qualitative Results Breakdown

Sample Code	Sample Content	Matrix	Sample Relationships	Detection Frequency	Sample Status	Percentage Correct (All)	
						(%)	(n)
CVOP20S-02	Coronavirus NL63	TM	4.64 dPCR Log10 Copies/ml	Negative	EDUCATIONAL	96.9	521




Individual Report	QCMD 2020 Coronavirus Outbreak Preparedness EQA Pilot Study				 <small>Quality Control for Molecular Diagnostics</small>	
	Catalogue Code: QAV204214	Ref Code: CVOP20	Challenge: S	Analysis Type: Qualitative	Dataset: 348360	Report UID: 11181/348360/2667


Groups below n=5: AB Analytica (n=3), AB Analytica - AB Analytica REALQUALITY RQ (n=3), ABL (n=1), ABL - ABL UltraGene (n=1), Anchor Diagnostics (n=3), Anchor Diagnostics - Anchor Diagnostics SARS CoV-2 (n=3), Applied Biosystems (n=2), Applied Biosystems - Applied Biosystems Real Time PCR Kit (n=2), Appolon biotek (n=1), Appolon biotek - Appolon biotek SARS-CoV-2 (n=1), AusDiagnostics - AusDiagnostics Easy-Plex (n=4), BGI (n=2), BGI - BGI Real Time PCR (n=2), BioGX (n=1), BioGX - BioGX SARS-CoV-2 (n=1), Bioeksen (n=2), Bioeksen - Bioeksen Biospeedy SARS-CoV-2 (n=2), CDC (n=3), CDC - CDC SARS-CoV-2 (n=3), Co-Diagnostics Inc (n=3), Co-Diagnostics Inc - Co-Diagnostics Inc Logix Smart (n=3), DiaCarta (n=1), DiaCarta - DiaCarta QuantiVirus PCR kit (n=1), DiaSorin (n=2), DiaSorin - DiaSorin Simplexa (n=2), Diagenode (n=1), Diagenode - Diagenode Real-Time PCR (n=1), EUROIMMUN (n=3), EUROIMMUN - EUROIMMUN SARS-CoV-2 (n=3), GenMark (n=1), GenMark - GenMark SARS-CoV-2 (n=1), GeneProof (n=4), GeneProof - GeneProof SARS-CoV-2 (n=4), Genematrix (n=1), Genematrix - Genematrix NeoPlex (n=1), Hologic (n=2), Hologic - Hologic Panther Fusion (n=2), HybriBio Limited (n=1), HybriBio Limited - HybriBio Real-Time PCR Kit (n=1), Immundiagnostik (n=1), Immundiagnostik - Immundiagnostik MutaPLEX (n=1), Kogene Biotech (n=1), Kogene Biotech - Kogene Powerchek (n=1), Liferiver (n=1), Liferiver - Liferiver 2019-nCoV (n=1), Luminex - Luminex ARIES (n=2), Luminex - Luminex xTAG (n=4), Mediven (n=1), Mediven - Mediven GenoAmp PCR (n=1), Mikrogen (n=4), Mikrogen - Mikrogen ampliCube (n=4), Mobidiag (n=4), Mobidiag - Mobidiag Amplidiag (n=4), QIAGEN (n=4), QIAGEN - Qiagen QIAstat-Dx (n=4), Randox (n=2), Randox - Randox COVID-19 Array (n=1), Randox - Randox Extended Array (n=1), Sansure Biotech (n=1), Sansure Biotech - Sansure Biotech nCoV PCR Kit (n=1), SeroSep (n=1), SeroSep - SeroSep Real Time PCR (n=1), Shanghai BioGerm (n=1), Shanghai BioGerm - Shanghai BioGerm Real Time PCR (n=1), Shanghai Fosun Pharmaceutical (n=1), Shanghai Fosun Pharmaceutical - Shanghai Fosun Real Time PCR (n=1), Vazyme (n=1), Vazyme - Vazyme SARS-CoV-2 (n=1), Vitassay (n=4), Vitassay - Vitassay SARS-CoV-2 (n=4), bioMérieux (n=4), bioMérieux - BioFire FilmArray (n=1), bioMérieux - bioMérieux R-gene SARS-CoV-2 (n=3), fast-track DIAGNOSTICS (n=2), fast-track DIAGNOSTICS - FTD real time PCR (n=2)

Groups Rolled Up: Abbott - Abbott SARS-CoV-2 (n=17), Altona Diagnostics - Altona Diagnostics RealStar (n=28), Anatolia Geneworks - Anatolia Geneworks Bosphore nCoV (n=9), Cepheid - Cepheid Xpert SARS-CoV-2 (n=51), ELITech Group - Elitech GeneFinder (n=25), PrimerDesign - PrimerDesign Genesig (n=7), R-Biopharm - R-Biopharm RIDA Gene (n=5), Roche - Roche Cobas SARS-CoV 2 (n=37), TIB MOLBIOL - TIB MOLBIOL LightMix (n=27), ThermoFisher - ThermoFisher TaqPath COVID-19 (n=6), In-House - Real-time In-House PCR (n=119)

CVOP20S-04 - Qualitative Results Breakdown

Sample Code	Sample Content	Matrix	Sample Relationships	Detection Frequency	Sample Status	Percentage Correct (All)	
						(%)	(n)
CVOP20S-04	Coronavirus OC43	TM	4.03 dPCR Log10 Copies/ml	Negative	EDUCATIONAL	97.1	521



Individual Report	QCMD 2020 Coronavirus Outbreak Preparedness EQA Pilot Study					
	Catalogue Code: QAV204214	Ref Code: CVOP20	Challenge: S	Analysis Type: Qualitative	Dataset: 348360	Report UID: 11181/348360/2667

Groups below n=5: AB Analytica (n=3), AB Analytica - AB Analytica REALQUALITY RQ (n=3), ABL (n=1), ABL - ABL UltraGene (n=1), Anchor Diagnostics (n=3), Anchor Diagnostics - Anchor Diagnostics SARS CoV-2 (n=3), Applied Biosystems (n=2), Applied Biosystems - Applied Biosystems Real Time PCR Kit (n=2), Appolon biotek (n=1), Appolon biotek - Appolon biotek SARS-CoV-2 (n=1), AusDiagnostics - AusDiagnostics Easy-Plex (n=4), BGI (n=2), BGI - BGI Real Time PCR (n=2), BioGX (n=1), BioGX - BioGX SARS-CoV-2 (n=1), Bioeksen (n=2), Bioeksen - Bioeksen Biospeedy SARS-CoV-2 (n=2), CDC (n=3), CDC - CDC SARS-CoV-2 (n=3), Co-Diagnostics Inc (n=3), Co-Diagnostics Inc - Co-Diagnostics Inc Logix Smart (n=3), DiaCarta (n=1), DiaCarta - DiaCarta QuantiVirus PCR kit (n=1), DiaSorin (n=2), DiaSorin - DiaSorin Simplexa (n=2), Diagenode (n=1), Diagenode - Diagenode Real-Time PCR (n=1), EUROIMMUN (n=3), EUROIMMUN - EUROIMMUN SARS-CoV-2 (n=3), GenMark (n=1), GenMark - GenMark SARS-CoV-2 (n=1), GeneProof (n=4), GeneProof - GeneProof SARS-CoV-2 (n=4), Genematrix (n=1), Genematrix - Genematrix NeoPlex (n=1), Hologic (n=2), Hologic - Hologic Panther Fusion (n=2), HybriBio Limited (n=1), HybriBio Limited - HybriBio Real-Time PCR Kit (n=1), Immundiagnostik (n=1), Immundiagnostik - Immundiagnostik MutaPLEX (n=1), Kogene Biotech (n=1), Kogene Biotech - Kogene Powerchek (n=1), Liferiver (n=1), Liferiver - Liferiver 2019-nCoV (n=1), Luminex - Luminex ARIES (n=2), Luminex - Luminex xTAG (n=4), Mediven (n=1), Mediven - Mediven GenoAmp PCR (n=1), Mikrogen (n=4), Mikrogen - Mikrogen ampliCube (n=4), Mobidiag (n=4), Mobidiag - Mobidiag Amplidiag (n=4), QIAGEN (n=4), QIAGEN - Qiagen QIAstat-Dx (n=4), Randox (n=2), Randox - Randox COVID-19 Array (n=1), Randox - Randox Extended Array (n=1), Sansure Biotech (n=1), Sansure Biotech - Sansure Biotech nCoV PCR Kit (n=1), Serosep (n=1), Serosep - Serosep Real Time PCR (n=1), Shanghai BioGerm (n=1), Shanghai BioGerm - Shanghai BioGerm Real Time PCR (n=1), Shanghai Fosun Pharmaceutical (n=1), Shanghai Fosun Pharmaceutical - Shanghai Fosun Real Time PCR (n=1), Vazyme (n=1), Vazyme - Vazyme SARS-CoV-2 (n=1), Vitassay (n=4), Vitassay - Vitassay SARS-CoV-2 (n=4), bioMerieux (n=4), bioMerieux - BioFire FilmArray (n=1), bioMerieux - bioMerieux R-gene SARS-CoV-2 (n=3), fast-track DIAGNOSTICS (n=2), fast-track DIAGNOSTICS - FTD real time PCR (n=2)

Groups Rolled Up: Abbott - Abbott SARS-CoV-2 (n=17), Altona Diagnostics - Altona Diagnostics RealStar (n=28), Anatolia Geneworks - Anatolia Geneworks Bosphore nCoV (n=9), Cepheid - Cepheid Xpert SARS-CoV-2 (n=51), ELITech Group - Elitech GeneFinder (n=25), PrimerDesign - PrimerDesign Genesig (n=7), R-Biopharm - R-Biopharm RIDA Gene (n=5), Roche - Roche Cobas SARS-CoV 2 (n=37), TIB MOLBIOL - TIB MOLBIOL LightMix (n=27), ThermoFisher - ThermoFisher TaqPath COVID-19 (n=6), In-House - Real-time In-House PCR (n=119)

QCMD © 2020. The QCMD EQA programme samples, associated reports and data generated during this programme are intended for External Quality Assessment (EQA) and Proficiency Testing (PT) purposes only. QCMD operates according to a strict Code of Practice which is in line with ISO/IEC 17043 and associated standards. Data reported in QCMD programmes is representative of a laboratory's standard diagnostic testing protocols irrespective of the technology they use. The data provided in the reports are based on technical information provided by the individual laboratories as part of the assessment process, as such it does not constitute a formal technology method comparison. All text and images produced by QCMD are the property of QCMD unless otherwise stated.

The reproduction and use of these materials is not permitted without the express written consent of QCMD. The use of the information provided in QCMD reports for commercial purposes is strictly prohibited.